

# Ionizing radiation risks of cardiac imaging: estimates of the immeasurable

Philipp A. Kaufmann<sup>1,2\*</sup> and Juhani Knuuti<sup>3</sup>

<sup>1</sup>Cardiac Imaging, University Hospital Zurich, Zurich CH-8091, Switzerland; <sup>2</sup>Zurich Center of Integrative Human Physiology (ZIHP), Zurich, Switzerland; and <sup>3</sup>Turku PET Centre, Turku, Finland

Received 7 June 2010; revised 1 August 2010; accepted 3 August 2010; online publish-ahead-of-print 9 September 2010

This paper was guest edited by Professor Jeroen Bax, Leiden University Medical Center, University of Leiden, The Netherlands

For many decades, the search for a non-invasive visualization of the coronary arteries seemed to remain an unfulfilled promise to clinical cardiologists. Owing to the rapid refinements successfully implemented in computed tomography (CT) technology over the past few years, non-invasive imaging of coronary arteries is now not only feasible but also has become a reality in daily routine. This may—at least in part—have contributed to the fact that the number of CT scans performed in the USA has quadrupled since 1993.<sup>1</sup> Although in a recent US survey CT and nuclear imaging accounted for just 21% of the total number of procedures, they resulted in >75% of the total cumulative effective radiation dose. We have witnessed an impressive six-fold increase in the radiation dose from medical imaging delivered per patient over the last 3 decades.<sup>1–2</sup> Interestingly, half of all nuclear medicine procedures worldwide and 25% of all X-ray studies are performed in the USA (constituting 5% of the world's population), doubling and tripling that of other developed countries.<sup>3</sup>

In this context, it appears appropriate that the radiation exposure experienced by patients undergoing any medical imaging procedure has recently obtained a growing attention and publicity.<sup>4</sup> Although some surveys have investigated on the overall amount of radiation exposure (Table 1) from any medical imaging procedure,<sup>5</sup> others have focused specifically on the radiation dose to patients from cardiac imaging.<sup>6–7</sup> Among these, CT coronary angiography has faced the greatest attention, probably because this modern development has been introduced as a last cardiac imaging technique and also because CT is generally perceived as being associated with a high radiation dose to the patient. In fact, in its infancies, radiation doses >20 mSv were reported for a CT coronary angiography.<sup>8</sup> Although comparable doses have also been reported from some surveys for purely diagnostic coronary catheterization,<sup>6</sup> which is invasive and achieves only low diagnostic yield in the actual daily clinical routine,<sup>9</sup> this

has fuelled a vivid discussion on the potential harms arising from non-invasive CT coronary angiography, questioning the justification of its use in large populations and calling for more efficient radiation protection measures of patients undergoing CT angiography.<sup>8</sup>

Remarkably, whereas the potential benefits of medical imaging procedures are generally left unmentioned in the radiation safety discussion although they can be scientifically quantified, the risk of cancer from low radiation doses used in medical imaging can only be roughly estimated by statistical calculations based on assumptions of the linear no-threshold theory.<sup>4</sup> This means that data from Hiroshima are extrapolated down to the lowest doses, although no studies have ever verified the assumptions about cancer associated with the doses used in medical imaging. Instead, even the authors of the largest recent survey on low-dose ionizing radiation from medical imaging procedures<sup>5</sup> have agreed that the data associating low-dose radiation to cancer risk are not definitive.<sup>10</sup> Similarly, the Health Physics Society has concluded in a position statement that although there is substantial and convincing scientific evidence for health risks following high-dose exposures, risks of health effects for doses <50–100 mSv are 'either too small to be observed or are nonexistent'.<sup>11</sup>

Nevertheless, following the principle of keeping radiation exposure as low as reasonably achievable, several strategies to reduce radiation dose in CT coronary angiography have been explored, such as automated exposure control, electrocardiographically controlled tube modulation, and reduced tube voltage (from 120 to 100 kV) in non-obese patients.<sup>12</sup> A prospective controlled multicentre trial has confirmed that introduction of a collaborative radiation dose-reduction programme was associated with a 53% reduction in radiation dose from 21 to 10 mSv in patients undergoing CT coronary angiography.<sup>13</sup> A recent milestone in

The opinions expressed in this article are not necessarily those of the Editors of the *European Heart Journal* or of the European Society of Cardiology.

\* Corresponding author. Tel: +41 44 255 4196, Fax: +41 44 255 4414, Email: pak@usz.ch

Published on behalf of the European Society of Cardiology. All rights reserved. © The Author 2010. For permissions please email: journals.permissions@oup.com.

**Table 1** Effective dose estimates for selected studies and sources

Examination	Representative effective dose value (mSv)
Chest X-ray postero-anterior	0.02
Round-trip flight (Europe—North America)	
Average	0.05
Peaks during solar cosmic events <sup>20</sup>	>2
Smoking cigarettes (natural Po-210 in tobacco) <sup>21</sup>	2.8/year
Radon in home	
Average	3.2/year
Peak exposures <sup>22</sup>	>200/year
Coronary calcium scoring (prospective triggering)	1
CT coronary angiogram (64-slice)	
Without tube current modulation	20
With tube current modulation	12
Prospective triggering	2
Prospective triggering with high-pitch spiral	1
CT chest	7
CT abdominal	8
Diagnostic invasive coronary angiogram	7
Myocardial perfusion study	
Thallium stress/rest	25
Sestamibi (1-day) stress/rest	10
N-13 ammonia stress/rest	3

dose reduction has been achieved by introducing prospective ECG triggering, limiting scanning to a narrow pre-defined end-diastolic phase, which resulted in a massive 90% reduction in radiation dose down to an average of about 2 mSv without loss of image quality<sup>14</sup> or accuracy.<sup>15</sup> Very recent introduction of prospective high-pitch spiral scanning has enabled to lower radiation dose <1 mSv.<sup>16</sup> Similarly, substantial dose reduction in nuclear myocardial perfusion imaging to <2 mSv has been achieved by new reconstruction algorithms,<sup>17</sup> by introducing low-dose stress-only protocols suitable for hybrid imaging with low-dose CT coronary angiography,<sup>18</sup> and by implementing semiconductor detectors into latest generation gamma cameras allowing massive scan shortening or dose reduction.<sup>19</sup>

As CT coronary angiographies can now be achieved with a radiation dose <1 mSv,<sup>16</sup> the estimated risk of inducing a fatal malignancy (Table 2) is now in the range of the lifetime odds of dying from a lightning strike.<sup>7</sup> Thus, although we agree that the time is right to initiate long-term observational studies involving patients who have undergone imaging, we should at the same time stop making assumptions—invoked by health care professionals and

**Table 2** Estimated risks of fatal malignancy from radiation exposure and lifetime odds of dying due to selected underlying causes

Exposure	Lifetime odds of dying (per 1000 individuals)
Effective radiation dose	
1 mSv	0.05
10 mSv	0.5
20 mSv (yearly radiation worker allowance)	1
100 mSv (definition of low exposure)	5
Natural fatal cancer <sup>7</sup>	212
Motor vehicle accident <sup>7</sup>	11.9
Lightning strike <sup>7</sup>	0.013

the media—that are not adequately supported by data but may harm our patients by deferring them from a needed diagnostic procedure.

## Funding

P.A.K. was funded by a grant of the Swiss National Science Foundation (SFN grant No. 320030\_127604/1).

**Conflict of interest:** none declared.

## References

- Brenner DJ, Hall EJ. Computed tomography—an increasing source of radiation exposure. *N Engl J Med* 2007;**357**:2277–2284.
- National Council on Radiation Protection and Measurements. Ionizing radiation exposure of the population of the United States: recommendation of the National Council on Radiation Protection and Measurements. Report No. 160, Bethesda, MD; NCRP, March 2009.
- Beller GA. Importance of consideration of radiation doses from cardiac imaging procedures and risks of cancer. *J Nucl Cardiol* 2010;**17**:1–3.
- Lauer MS. Elements of danger—the case of medical imaging. *N Engl J Med* 2009;**361**:841–843.
- Fazel R, Krumholz HM, Wang Y, Ross JS, Chen J, Ting HH, Shah ND, Nasir K, Einstein AJ, Nallamothu BK. Exposure to low-dose ionizing radiation from medical imaging procedures. *N Engl J Med* 2009;**361**:849–857.
- Einstein AJ, Moser KW, Thompson RC, Cerqueira MD, Henzlova MJ. Radiation dose to patients from cardiac diagnostic imaging. *Circulation* 2007;**116**:1290–1305.
- Gerber TC, Carr JJ, Arai AE, Dixon RL, Ferrari VA, Gomes AS, Heller GV, McCollough CH, McNitt-Gray MF, Mettler FA, Mieres JH, Morin RL, Yester MV. Ionizing radiation in cardiac imaging: a science advisory from the American Heart Association Committee on Cardiac Imaging of the Council on Clinical Cardiology and Committee on Cardiovascular Imaging and Intervention of the Council on Cardiovascular Radiology and Intervention. *Circulation* 2009;**119**:1056–1065.
- Mollet NR, Cademartiri F, van Mieghem CA, Runza G, McFadden EP, Baks T, Serruys PW, Krestin GP, de Feyter PJ. High-resolution spiral computed tomography coronary angiography in patients referred for diagnostic conventional coronary angiography. *Circulation* 2005;**112**:2318–2323.
- Patel MR, Peterson ED, Dai D, Brennan JM, Redberg RF, Anderson HV, Brindis RG, Douglas PS. Low diagnostic yield of elective coronary angiography. *N Engl J Med* 2010;**362**:886–895.
- Fazel R, Krumholz HM, Nallamothu B. Radiation exposure from medical imaging procedures; author reply. *N Engl J Med* 2009;**23**:2291–2292.
- Health Physics Society. Radiation risk in perspective. Position statement of the Health Physics Society. 2010. PS010-2. <http://hps.org/hpspublications/positionstatements.html> (13 August 2010).
- Hausleiter J, Meyer T, Hermann F, Hadamitzky M, Krebs M, Gerber TC, McCollough C, Martinoff S, Kastrati A, Schomig A, Achenbach S. Estimated radiation dose associated with cardiac CT angiography. *JAMA* 2009;**301**:500–507.

13. Raff GL, Chinnaiyan KM, Share DA, Goraya TY, Kazerooni EA, Moscucci M, Gentry RE, Abidov A. Radiation dose from cardiac computed tomography before and after implementation of radiation dose-reduction techniques. *JAMA* 2009;**301**:2340–2348.
14. Husmann L, Valenta I, Gaemperli O, Adda O, Treyer V, Wyss CA, Veit-Haibach P, Tatsugami F, von Schulthess GK, Kaufmann PA. Feasibility of low-dose coronary CT angiography: first experience with prospective ECG-gating. *Eur Heart J* 2008;**29**:191–197.
15. Herzog BA, Husmann L, Burkhard N, Gaemperli O, Valenta I, Tatsugami F, Wyss CA, Landmesser U, Kaufmann PA. Accuracy of low-dose computed tomography coronary angiography using prospective electrocardiogram-triggering: first clinical experience. *Eur Heart J* 2008;**29**:3037–3042.
16. Achenbach S, Marwan M, Ropers D, Schepis T, Pflederer T, Anders K, Kuettner A, Daniel WG, Uder M, Lell MM. Coronary computed tomography angiography with a consistent dose below 1 mSv using prospectively electrocardiogram-triggered high-pitch spiral acquisition. *Eur Heart J* 2010;**31**:340–346.
17. Pazhenkottil AP, Herzog BA, Husmann L, Buechel RR, Burger IA, Valenta I, Landmesser U, Wyss CA, Kaufmann PA. Non-invasive assessment of coronary artery disease with ct coronary angiography and spect: a novel dose-saving fast-track algorithm. *Eur J Nucl Med Mol Imaging* 2010;**37**:522–527.
18. Husmann L, Herzog BA, Gaemperli O, Tatsugami F, Burkhard N, Valenta I, Veit-Haibach P, Wyss CA, Landmesser U, Kaufmann PA. Diagnostic accuracy of computed tomography coronary angiography and evaluation of stress-only single-photon emission computed tomography/computed tomography hybrid imaging: comparison of prospective electrocardiogram-triggering vs. retrospective gating. *Eur Heart J* 2009;**30**:600–607.
19. Herzog BA, Buechel RR, Katz R, Brueckner M, Husmann L, Burger IA, Pazhenkottil AP, Valenta I, Gaemperli O, Treyer V, Kaufmann PA. Nuclear myocardial perfusion imaging with a cadmium-zinc-telluride detector technique: optimized protocol for scan time reduction. *J Nucl Med* 2010;**51**:46–51.
20. Bütikofer R, Flückiger EO, Benoît P, Desorgher L. Effective radiation dose for selected intercontinental flights during the GLEs on 20 January 2005 and 13 December 2006. In *Proc. 2009; European Cosmic Ray Symposium* 1–5.
21. Thompson RC, Cullom SJ. Issues regarding radiation dosage of cardiac nuclear and radiography procedures. *J Nucl Cardiol* 2006;**13**:19–23.
22. Radioprotection and surveillance of radioprotection in Switzerland: 2009 Report. <http://www.bag.admin.ch/themen/strahlung/00043/00065/02238/index.html?lang=de> (25 August 2010).